

TO: Users and Distributors of TEGAM Model 110A Safety Voltmeters  
FROM: Kevin Kaufman, Director of National Sales  
DATE: January 1, 2017  
SHTS: 1  
SUBJ: Changes in UL 61010-1 2<sup>nd</sup> Edition that Effect TEGAM Model 110A Safety Voltmeters

---

Dear Valued Customer,

The TEGAM model 110A is currently listed by Underwriters Laboratories, (UL), to comply with UL 61010-1 2<sup>nd</sup> Edition. UL has announced that as of January 1, **2018**, it will no longer list any products under UL 61010-1 2<sup>nd</sup> Edition, and will require instruments to meet the new standard UL 61010-1 3<sup>rd</sup> Edition, where resistive probes, like those in the 110A, are no longer considered preferred safety devices.

To bring the 110A design into compliance with the new standard would require us to stop using resistive probes and completely redesign the instrument with some other redundant safety feature. Because we see the safety value of the 110A being directly tied to the existence of resistive probes, we believe that removing them would completely undermine the inherent safety of the 110A and eliminate its value to you, our customer.


Though we will continue to offer the 110A going forward, effective January 1, **2018**, the TEGAM model 110A will no longer carry the UL label. We struggled with this decision and are using this forum to report that we will not commit resources to a redesign effort at this time.

From our vantage we see two courses of action that our customers can take forward:

1. Purchase an inventory of UL Listed 110A that can be delivered before January 1, 2018 to provide a supply of instruments until a replacement can be identified. (For delivery in December we must receive the order before August 31, 2017.)
2. Decide to continue using the 110A regardless of its status with UL or CE.

These instruments will continue to be manufactured with the same design and quality control care that has been a hallmark of TEGAM since the product's original introduction. We look forward to your continued patronage and are available to discuss any questions you may have. Please feel free to contact me directly at 330-416-1232 or [kkaufman@tegam.com](mailto:kkaufman@tegam.com).

Thank you and best success,



T. Kevin Kaufman  
Director of National Sales